Coated metal surface on solid support for displacement reactions.

The present invention relates to a coated metal surface on solid support for displacement reactions, especially for analyte detection in an aqueous solution by displacement from the metal surface coating of reversibly bound antibodies specific for the analyte. Detection of the

5   displacement, and thus the presence of the analyte in an aqueous solution is performed with a analysis device, such as a Piezoelectric Crystal Microbalance (PCM)device or a Surface Plasmon Resonance (SPR) biosensor.

Background

The SPR biosensor is a sensitive real-time technique, which can be used to

10   extract information about molecular interaction near certain metal surfaces. It offers the possibility to determine concentration, association and dissociation rate constants and affinity as well as epitope mapping and determination of interaction specificity [B. Liedberg and K. Johansen, *Affinity biosensing based on surface plasmon detection* in "Methods in Biotechnology, Vol. 7: Affinity Biosensors: Techniques and Protocols", K. R. Rogers and A.

15   Muchandani (Eds.), Humana Press Inc., Totowa, NJ, pp. 31-53]. One of the components participating in the studied reaction is immobilized on the metal surface either before or during the SPR experiment. The immobilized molecule is exposed to a continuous flow into which one can inject interacting species. The method is based on optical detection and the sensing signal reflects changes in dielectric function or refractive index at the surface. These

20   changes can be caused by molecular interaction at the surface.

The PCM technique is based on an oscillating piezoelectric crystal in a microbalance device, wherein the crystal consists of e.g. quartz, aluminum nitride (AlN) or sodium potassium niobiates (NKN). When the crystal is a quarts crystal, the device is referred to as a QCM (quartz crystal microbalance). The PCM and QCM are gravimetrical sensors and

25   are thus sensitive to mass changes. A QCM comprises a piezoelectric quartz crystal plate upon which metal electrodes have been deposited on both sides. An alternating potential difference applied on such a crystal plate induces shear waves. At certain frequencies – such that the thickness is an odd integer of half wavelengths – the crystal will be in resonance [M. Rodahl, F. Höök, A. Krozer, P. Brzezinski and B. Kasemo, *Quartz crystal microbalance setup*

30   *for frequency and Q-factor measurements in gaseous and liquid environments*, Review of Scientific Instruments 66 (1995) pp. 3924-3930]. The energetically most favourable number of half wavelengths is one. The resonance frequency is dependent on the thickness of the crystal, but is normally in the MHz range. A mass change on the surface of the plate will result in a shift in the resonance frequency. The fact that frequency shifts of 0.01 Hz can be

easily measured makes the QCM a sensitive sensor for determining mass variations. A number of patents and other publications describe the use of piezoelectric quartz crystals (QCM) as affinity-based chemical sensors/detectors in e.g. various immunoassay techniques, and detection of bacteria and virus. In most of these applications the QCM-instrument is used to analyze the weight gain of the crystal after interaction between interaction pairs, e.g. antibodies and antigens.

There are obvious difficulties in analyzing small molecules with conventional immunosensors due to the low response, i.e. small change in weight of the sensor crystal. For attaining the necessary detection of small molecules, the sensitivity of the system has to be improved. To improve the detectability of small molecules, they should be reacted with larger molecules by specific interaction between the small molecules and the larger ones. For example, small antigen molecules are reacted with antibodies specifically binding to them to form antigen-antibody complexes that are easier to detect. If an antigen derivative with less affinity to the antibody than the analyte antigen is immobilized to a surface, antibodies specific for these antigens may be reversibly bound to the immobilized antigen. Then, when the analyte antigen is present in a solution, the antibody will be displaced from the immobilized antigen and form an antigen-antibody complex with the analyte antigen. In case the antibody carries a marker, such as a fluorescent label, the formed complex can be detected with the aid of the marker. On the other hand, if the immobilized antigen is immobilized on a surface of a biosensor sensitive for mass changes, then the displacement of the antibody from the surface will result in a weight loss. Such displacement reactions are used in the present invention.

An organosulphur compound, such as an alkyl thiol, can be used to form a well-ordered and densely packed SAM on a gold substrate. The strong chemical bond between sulphur and gold couples the molecules to the surface. Once pinned to the substrate, which occurs within seconds, the molecules start to organize themselves into densely packed formations due to the van der Waal forces between the alkyl chains. The latter process is time consuming and it takes hours or even days before a well-ordered SAM is completed. The length of the molecules used has a strong influence on the properties of the obtained SAM. An all trans SAM is a perfectly ordered and densely packed monolayer, whereas a SAM of poorer quality possesses defects like complete or terminal gauche (more or less spaghetti-like). The molecules in a SAM will display a chain tilt of 25-40° due to the mismatch between the pinning distance and size of the van der Waal diameter of the carbon chains [B. Liedberg and J. M. Cooper, *Bioanalytical applications of self-assembled monolayers* in "Immobilized

Biomolecules in Analysis: A Practical Approach", T. Cass and F. S. Liegler (Eds.), Biosensors, Oxford university press, Oxford, pp. 55-78]. The free end of the molecule can be linked to desired groups or even proteins. In this way it is possible to design surfaces with interesting and useful properties. Mixing different alkyl thiols further increases the versatility.

5   However, it should be noted that a certain mixture of thiols in a loading solution does not necessarily result in a SAM of the same mixture. On the contrary, this is seldom the case due to complex thermodynamic processes taking place during the self-assembly.

To our knowledge, SAM coupled to a low molecular weight antigen has not been developed for displacement reactions where an antigen-specific antibody is reversibly bound to the

10 immobilized antigen and dissociates in aqueous solution and binds to an analyte antigen that has a higher affinity to the antibody than the immobilized antigen.

Description of the invention

The present invention provides a coated metal surface on a solid support that is useful in an analysis device for detection of an analyte antigen in an aqueous solution by

15 monitoring displacement of an antibody reversibly bound to an antigen on the coating by dissociation and reaction with the analyte antigen.

In this specification and claims the word antibody is intended to comprise whole antibodies or antigen-binding parts of antibodies or synthetic antigen-binding molecules.

Thus, one aspect of the invention is directed to a coated metal surface on a solid

20 support, wherein the coating consists of a self-assembled monolayer (SAM) of oligo(ethylene glycol)-terminated amide group-containing alkyl (OEG) thiols. The OEG thiols contain a SH group that is firmly attached to the metal surface and a low molecular weight antigen introduced via amide-groups to the SAM-forming OEG thiol molecule, wherein the alkyl portion has 1 to 20 methylenes, the OEG portion has 1 to 15 ethylene oxy units, and wherein

25 the antigens are optionally reversibly bound to antibodies specific for the antigens.

In an embodiment of the invention, the oligo(ethylene glycol) has 4-6 ethylene oxy units and the alkyl group has 15 methylene units.

The low molecular weight antigens are synthetically bound to the OEG molecules prior to SAM formation by reacting functional groups on the antigens with

30 functional groups terminating the OEG thiol. These functional groups can be of the type-carboxylic acid, amino and hydroxyl groups. It may be necessary to introduce a functional group on the low molecular weight antigen prior to the reaction in case the antigen lacks functional groups for the reaction.

4

The coated metal surface on a solid support according to the invention will usually be stored separately from the antigen-specific antibodies prior to use. When used in displacement analysis, the coated metal surface on a solid support will, however, comprise the specific antibodies reversibly bound to the antigens of the coating.

5   The metal of the coated metal surface on a solid support according to the invention is preferably selected from e.g. the group consisting of gold, silver, aluminum, chromium and titanium. The presently preferred metal is gold.

The antigen of the coating is the same as or a derivative of the analyte antigen except that it is immobilized through a bond to the SAM. The antigen of the coating may thus be derivatized

10   to optimize dissociation of the bound antibody in an aqueous solution.

The antigens bound to the SAM of the coating according to the invention are the same or different, i.e. the antigens may bind to the same specific antibodies or there may be a mixture of two or more bound antigens binding to different specific antibodies enabling the detection of the presence of several different analyte antigens in an aqueous solution. In case

15   the antibodies carry different markers, such as fluorescent markers, it will be possible to detect displacement of the different antibodies. However, a mixture of several different antibodies will normally be used in cases where screening of samples for any of the target antigens is sufficient, such as screening of samples for any narcotics or explosives. In order to avoid interferences between the different antigens and antibodies, having different affinities to

20   each other, it may be necessary to introduce discrete patches or microarrays of spots of coatings with the different antigens on the solid support.

In a preferred embodiment of the invention the antigen of the coating is selected from the group consisting of optionally derivatized explosives and narcotics. In case the selected antigen of the coating binds too strongly to the specific antibody so that the dissociation of the

25   antibody in aqueous solution is hampered, the antigen molecule may be chemically modified, e.g. by modification of functional groups such as ester or amino groups (by removal of, or replacing the original groups) or by eliminating a part of the antigen molecule, or introducing new functional groups or side chains to the antigen molecule, to reduce its affinity to the antibody.

30   The explosives are preferably selected from the group consisting of trinitrotoluene (TNT), dinitrotoluene (DNT), hexahydro-1,3,5-trinitro-1,3,5-triazine (RDX), octahydro-1,3,5,7-tetranitro-1,3,5,7-tetrazine (HMX), pentaerythritol tetranitrate (PETN), and nitroglycerine (NG), and the narcotics are preferably selected from the group consisting of

cocaine, heroine, amphetamine, methamphetamine, cannabinols, tetrahydrocannabinols (THC), and methylenedioxy-N-methylamphetamine (Ecstacy).

In a presently preferred embodiment the solid support of the coated metal surface on a solid support according to the invention is a piezoelectric quarts crystal electrode or a glass plate or prism, and thus the coated metal surface on the piezoelectric crystal electrode is suitable for use in a PCM device and the coated metal surface on a glass plate or prism is suitable for use in a SPR apparatus.

Another aspect of the invention is directed to the use of the coated metal surface on a solid support according to the invention as part of an analysis device for detection in an aqueous solution of analyte antigens with higher affinity to specific antibodies than the antigens of the coating by monitoring the displacement of the antibodies from the coating.

Yet another aspect of the invention is directed to a method of detecting analyte antigens in an aqueous solution comprising activating, if necessary, the coated metal surface on a solid support according to the invention lacking bound antibodies by bringing antigen-specific antibodies into contact with the coated metal surface in an aqueous solution, allowing binding of the antibodies to the antigens of the coating, removing excess antibodies, bringing the aqueous solution possibly containing the analyte antigens that have higher affinity to the antibodies than the antigens of the coating into contact with the antibodies reversibly bound to the coating, allowing the antibodies to dissociate and react with the analyte antigens, and detecting the loss of mass on the coated metal surface by means of an analysis device.

In an embodiment of the method of the invention the analysis device is selected from the group consisting of a Piezoelectric Quarts Crystal Microbalance device and a Surface Plasmon Resonance biosensor.

In a presently preferred embodiment the analysis device comprises a flow cell in which the coated metal surface on a solid support according to the invention is placed.

The invention will now be illustrated by some drawings and description of experiments, but it should be understood that the invention is not intended to be limited to the specifically described details.

Description of the drawings

Fig.1 shows the chemical formula of some narcotics, Cocaine, Heroine, Amphetamine, Ecstacy, Methamphetamine ,Cannabinol and Tetrahydrocannabinol (THC).

Fig. 2 shows the chemical formula of trinitrotoluene (TNT) and 2,4-dinitrotoluene (2,4-DNT).

Fig. 3 shows schematically the displacement mechanism taking place on the metal surface of a solid support, e.g. a QCM electrode. Note that the representation is not to scale. In reality an antibody is much larger than a TNT molecule. The sensor surface displayed is based on the self-assembly (SAM) technique. The derivatized TNT molecules,

5   TNT-analogs, are covalently bound to the metal surface via the SAM, the ABTNT antibodies specific for the TNT and TNT-analog is at first reversibly, loosely, bound to the TNT-analogs, and at exposure to TNT in solution, the ABTNT dissociates and forms a complex with TNT.

Fig. 4 shows the chemical structures of EG_4 and EG_6.

10   Fig. 5 shows the chemical structure of ANA1, ANA2 and ANA3.

Fig. 6 shows adsorption of antibodies, after 30min incubation in ABTNT (0.02g/L), to different TNT-analogues at different mixing ratios with EG_4. An excellent agreement is observed between the two techniques, IRAS and null ellipsometry.

Fig. 7 shows the ABTNT-binding capacity observed for EG_4 and ANA1, and

15   mixtures thereof, by the real-time technique SPR (Biacore2000). The flow rate was set to 50μL/min and a volume of 100μL of ABTNT (0.02g/L) was injected. Note the low adsorption onto the SAM of EG_4.

Fig. 8 shows ABTNT-binding capacity observed for EG_4 and ANA1, and mixtures thereof, by the real-time technique QCM (V2B). The flow rate was set to 50μL/min

20   and a volume of 100μL of ABTNT (0.02g/L) was injected. Note the low adsorption onto the SAM of EG_4.

Fig. 9 shows experiments on a Biacore2000 instrument showing SPR response to TNT injections of 1, 10 and 100pg/μL for SAMs of EG_4 and ANA1, and mixtures thereof. The surfaces had previously been loaded by injections of 100μL ABTNT. The flow rate was

25   set to 50μL/min.

Fig. 10 shows QCM measurements performed on a modified V2B flow cell system. The flow was set to 50μL/min and all TNT injection volumes were 100μL. The surfaces had previously been loaded with ABTNT (0.02g/L) by an injection of 100μL. The SAMs were assembled from loading solutions containing 100% and 50% ANA1. The

30   concentrations of TNT in the injections were 1, 10 and 100pg/μL and they were made in series, leaving the surface previously exposed to TNT for the second and third TNT injection. The arrows show the injections.

Fig. 11 shows QCM measurements performed on a modified V2B flow cell system. The flow was set to 50μL/min and all TNT injection volumes were 100μL. The

surfaces had previously been loaded with ABTNT (0.02g/L) by an injection of 100μL. The
SAMs were assembled from loading solutions containing 10% and 1% ANA1. The
concentrations of TNT in the injections were 1, 10 and 100pg/μL and they were made in
series, leaving the surface previously exposed to TNT for the second and third TNT injection.
The arrows show the injections.

Description of experiments

A mixed monolayer was produced that contained two kinds of molecules, the
first being protein repellent and the second being a TNT-analogue, thereby making it possible
to obtain SAMs containing a varying amount of analogue that displays low levels of non-
specific binding.

The initial step was to evaluate the protein resistant properties of a SAM
constituted of oligo(ethylene glycol) (OEG)-terminated amide group-containing alkyl thiols.
The two molecules chosen for this purpose were EG_4 and EG_6 (Fig 4). Previous reports have
shown repellent qualities for these molecules [P. Harder, M. Grunze, R. Dahint, G. M.
Whitesides and P. E. Laibinis, Molecular conformation in oligo(ethylene glycol)-terminated
self-assembled monolayers on gold and silver surfaces determines their ability to resist
protein adsorption, Journal of Physical Chemistry B, 102 (1998) pp. 426-436]. SAMs of pure
EG_4 and EG_6, respectively, as well as different compositions of the two were characterized by
use of several techniques, namely null ellipsometry, contact angle goniometry and infrared
reflection absorption spectroscopy (IRAS).

Furthermore, three TNT-analogue molecules (ANA1, ANA2 and ANA3)
(Figure 5), all containing a 2,4-dinitrobenzene end group, were examined separately and in
different mixings with a suitable candidate among the OEG compositions mentioned above.
These mixed SAMs were then characterized by the same techniques used for the pure OEG
SAMs. In addition, SAMs made from 100% analogue – 1,2 and 3 separately – were examined
with X-ray photoelectron spectroscopy (XPS).

Using null ellipsometry and IRAS, the capacity to bind ABTNT was also
determined for all the TNT-analogues at the different mixing ratios. Moreover, the magnitude
of displacement of ABTNT in response to exposure to TNT was examined by means of SPR
and QCM.

<u>Compounds</u>

The EG_4 and EG_6 molecules were obtained from the divisions of Applied
Physics and Chemistry, Department of Physics and Measurement Technology, Linköping
University, Sweden and the analogues ANA1, ANA2 and ANA3 were synthesized at the

division of Chemistry, Department of Physics and Measurement Technology, Linköping
University, Sweden. ABTNT and TNT were from Biosensor Applications Sweden AB.

Sample preparation

Silicon wafers were cleaned in TL2 (MilliQ water:25% hydrogen peroxide:37%

5   hydrogen chloride 6:1:1 at 85°C for 10min) and rinsed thoroughly in MilliQ water and dried
in nitrogen gas prior to coating of 25Å of titanium and 2000Å of gold by electron beam
evaporation. The equipment used was a Balzers UMS 500 P system. The evaporation rate was
1Å/s and 10Å/s for titanium and gold, respectively. A base pressure of at least 10^{-9} was kept
and during evaporation the pressure was noted to be on the low 10^{-7} scale at all times. This

10   type of surfaces was used for all experiments except SPR and QCM measurements. The SPR
surfaces (plain gold) were obtained from Biacore AB, Uppsala, Sweden and the gold-coated
QCM crystals were obtained from Biosensor Applications Sweden. It should be noted that the
surfaces used for SPR experiments had a similar surface roughness compared to the ones used
for the rest of the experiments, whereas the surface coating of the QCM crystals were of a

15   much rougher nature.

Before exposure to thiol loading solutions the sample surfaces were cleaned in
TL1 (MilliQ water:25% hydrogen peroxide:30% ammonia 5:1:1 at 85°C for 10min) and
rinsed thoroughly in MilliQ water. The concentration of thiols in the 99.5% ethanol based
loading solutions was 20μM for pure thiol solutions as well as for mixed thiol solutions.

20   Incubation of the surfaces occurred during approximately 40h at room conditions. The
samples were rinsed in 99.5% ethanol twice and then ultrasonicated for 3min (since the gold
coating on the QCM crystals did not withstand this step, it was omitted) and rinsed two more
times in 99.5% ethanol. If not stated otherwise, the surfaces were stored in pure 99.5%
ethanol for a maximum of 8h before they were dried in nitrogen gas and analysed. A number

25   of samples examined with IRAS and null ellipsometry were also subsequently incubated at
room conditions for 30min in ABTNT at a concentration of 0.02g/L, prepared in PBS (pH7.4)
and examined again. At all time the samples were handled with TL1-cleaned forceps.

Surface plasmon resonance (SPR)

For SPR experiments two different types of instruments from Biacore AB,

30   equipped with temperature controlled flow cells, were employed. A series of experiments
were performed on a BiacoreX apparatus with two flow channels. For these experiments the
flow rate was set to 10μL/min and the sample surfaces were loaded by a 70μL injection of
ABTNT (0.02g/L). Subsequently, TNT solutions of concentrations 100pg/μL and 10ng/μL

were injected in separate flow channels to cause displacement and hence a dissociation of ABTNT from the surface.

The second instrument used was a Biacore2000 equipped with four flow channels. The flow rate was 50μL/min and all injection volumes were 100μL. The four flow channels were run in series at all times. Like before, the injected ABTNT had the concentration 0.02g/L. The concentrations of the TNT injections were 1, 10 and 100pg/μL.

In all cases the running buffer was PBS (pH7.4) and both ABTNT and TNT solutions were prepared in the same medium. The sample surfaces used for SPR experiment were glass plates coated with about 400 Å gold and the flow cell temperature was kept at 25°C.

Quartz crystal microbalance (QCM)

The QCM measurements were performed at room conditions on a slightly modified flow cell system V2B from Biosensor Applications Sweden AB.. The AT-cut QCM crystals used were a thickness shear mode type with a resonance frequency of 10MHz. The thickness of the deposited titanium and gold layer were 250-300Å and 400-450Å, respectively. All parameters were set as in the experiments with Biacore2000, i.e. flow rate 50 μL/min, injection volume 100μL, ABTNT concentration 0.02g/L, TNT concentrations 1, 10 and 100pg/μL and running buffer PBS (pH7.4). It should be noted that the TNT injections were made one after another in the same flow channel, which means that only the 1pg/μL TNT injection was in fact performed on a TNT-non-exposed surface.

Results

OEG molecules

The OEG molecules EG_4 and EG_6 were used to produce SAMs on gold. Besides pure EG_4 and EG_6 SAMs, two mixed SAMs containing both molecules, were prepared and examined. The latter ones were assembled from loading solutions containing 75% and 50% of EG_4 and the rest EG_6. The mixed monolayers were denoted EG_4:EG_6 3:1 and EG_4:EG_6 1:1. The SAM characterization of different compositions of EG_4 and EG_6 with null ellipsometry and contact angle goniometry are given in Table 1. The self-assembly process displays good repeatability and the obtained results agree with recent findings [R. Valiokas, M. Östblom, S. Svedhem, S. C. T. Svensson, and B. Liedberg; *Temperature-driven phase transitions in oligo(ethylene glycol)-terminated self-assembled monolayers*, The Journal of Physical Chemistry B, 104(32) (2000) pp. 7565-7569, REF.]. The small angles obtained in contact angle goniometry measurements display the low hydrophobicity of these SAMs, which is one

of the many prerequisites for repellent properties. A slight increase in thickness can be discerned with increasing proportion of EG_6. As a rule of thumb, hydrophobic surfaces attract proteins and cells. Furthermore, the low value of the hysteresis suggests very homogenous surfaces.

5

Table 1. Characteristics of SAMs of EG_4, EG_6 and mixtures of the two, given with maximum errors.

Loading solution	Thickness of SAM (Å)		Contact angle goniometry (°)			
			Advancing θ_a		Receeding θ_r	
EG_4	35.7 ± 2.2[a]	(33.9^e)	29 ± 3[c]	(30^e)	24 ± 1[c]	(28^e)
$EG_4:EG_6$ 3:1	34.4 ± 0.5[b]		31 ± 1[d]		27 ± 1[d]	
$EG_4:EG_6$ 1:1	37.0 ± 1.5[a]		33 ± 1[c]		30 ± 1[c]	
EG_6	38.4 ± 0.7[b]	(38.9^e)	28 ± 3[d]	(28^e)	23 ± 1[d]	(25^e)

[a] 3×5 measurements on three surfaces, [b] 2×5 measurements on two surfaces,

[c] three measurements on three surfaces [d] two measurements on two surfaces,

10   [e] REF Valiokas et al..

TNT-analogues

Before the TNT- analogues were mixed with EG_4, they were examined separately in SAMs assembled from loading solutions containing pure ANA1, 2 and 3,

15   respectively. The results from ellipsometry and contact angle measurements are summarized in Table 2. As expected, the thickness of the analogues exceeds that of the OEG molecules. The small differences between the three analogues even reflect the difference in length of the molecules, recalling their chemical structure in Figure 5. The contact angles noted here are generally larger than those seen for the OEG molecules. Also the hysteresis between

20   advancing and receeding angle is larger and reveals a rougher surface. This would be expected, considering the relatively bulky dinitrobenzene end groups facing away from the surface, which could introduce defects in the produced SAMs.

25

Table 2. Characteristics of SAMs of ANA1, ANA2 and ANA3, given with maximum errors.

Loading solution	Thickness of SAM (Å)	Contact angle goniometry (°)	
		Advancing θ_a	Receeding θ_r
ANA1	47.0 ± 0.9^a	66 ± 2^b	52 ± 1^b
ANA2	48.4 ± 0.7^a	53 ± 2^b	38 ± 1^b
ANA3	49.8 ± 0.5^a	51 ± 3^b	34 ± 1^b

Immobilization of ABTNT

The ability to bind ABTNT has been evaluated for the different TNT-analogues
in their different mixing ratios. Several techniques have been used and the agreement between
them is striking.

The two diagrams, illustrated in Figure 6, are based on IRAS and ellipsometric
measurements and show the amount of immobilized ABTNT for the different analogues and
their mixing ratios with EG_4. For the IRAS data, part of amide I band was used as a measure
of bound ABTNT, integrating between 1710-1665cm^{-1}. The ellipsometric data shows the
increase in film thickness after incubation in ABTNT. Here, EG_4 shows its protein repellent
properties. The amount of ABTNT immobilized is virtually zero. Furthermore, the binding of
the antibody is quite similar for the three SAMs containing high amounts of analogue. The
SAMs assembled from 1% analogue solutions generally displays a lower degree of
immobilization.

Functionality test

The functionality of the three TNT-analogues in their different mixing ratios has
been evaluated with two aspects in mind. First, their ABTNT-binding capacity has been
considered and second, the dissociation of ABTNT in response to TNT exposure. In both
cases the two real-time techniques, SPR and QCM, have been employed. For the SPR
measurements an increase in response units (RU) corresponds to an increased amount of
bound ABTNT on the surface, while for QCM experiments a frequency drop is the
equivalent. All three analogues possess a high potential, but focus has been on ANA1, since it
displayed a slightly better performance than the others. For all experiments in this section the
running buffer was PBS (pH7.4) and both ABTNT and TNT solutions were prepared in the
same medium. The ABTNT concentration was always 0.02g/L.

12

ABTNT-binding capacity

For the SPR experiments two types of instruments were used, namely a Biacore2000 and a BiacoreX system from Biacore AB. For the measurements performed on the Biacore2000 apparatus the flow rate was set to 50µL/min and the sample surfaces were loaded by a 100µL injection of ABTNT. In Figure 7 the ABTNT-binding capacity of EG$_4$ and ANA1, and mixtures thereof, is visualized. A very low adsorption is seen for the SAM of pure EG$_4$ further supporting its protein repellent properties. The same mutual relationship between the different mixing ratios was found in experiments with the BiacoreX instrument run at 10µL/min and ABTNT injection volumes of 70µL (data not shown).

The QCM measurements were performed on a slightly modified flow cell system V2B developed by Biosensor Applications Sweden AB. All parameters were set as for the Biacore2000 experiments, i.e. flow rate 50 µL/min and ABTNT injection volume 100µL. Curves, showing the ABTNT-binding capacity of EG$_4$ and ANA1 and mixtures thereof, are seen in Figure 8. Once again, the low ABTNT-binding capacity of EG$_4$ is clearly demonstrated.

The binding of ABTNT to the different SAMs is similar for SPR and QCM experiments. The three surfaces containing most TNT-analogue all bind ABTNT very well to the surface and the dissociation of the antibodies is very slow. A certain release of ABTNT is expected due to the constant exposure to fresh buffer, i.e. a true equilibrium can never be reached.

ABTNT displacement

SPR curves (Biacore2000, flow rate: 50µL/min) showing ABTNT desorption in response to TNT injection of 1, 10 and100pg/µL for the different mixing ratios of EG$_4$ and ANA1, are seen in Figure 9. The surfaces had previously been loaded by injections of 100µL ABTNT.

The appearance of the curves in Figure 9 indicates that ABTNT binds weaker to the SAMs containing less ANA1, thereby facilitating the displacement reaction. This could be a consequence of the bivalency of the antibodies and their interaction with the surface. The higher the content of ANA1 is, the greater the chance for an ABTNT to find two TNT-analogues, one for each epitope, to bind to. In the case of EG$_4$:ANA1 99:1 this event might be less likely to occur, simply because ANA1 is less abundant. Since the binding strength of an antibody is strongly dependent on whether only one or both of its epitopes have bound to antigens, this aspect is highly relevant.

In Figure 10 and 11 corresponding results from QCM measurements (flow rate: 50µL/min, injection volume: 100µL) are shown. The mass loss caused by the TNT injections is seen as an increase in resonance frequency. The derivative of the frequency df/dt, which is proportional to the concentration, is also included, usually providing a clearer detection signal.